



CERTIFICATE OF ANALYSIS

KIT NAME: Mythic 18-22 Enzymatic
Cleaning Solution
CAT. NO: HM18-007-1
LOT NO: 709-18
EXPIRY DATE: 2019-09
DATE OF QC: 25.09.2017

<i>PARAMETER</i>	<i>NORM</i>	<i>RESULTS</i>
Appearance	clear	clear
Colour	colourless	colourless
Odour	characteristic	characteristic
pH (25°C±0.5)	7.45 – 7.65	7.50
Conductivity (mS/cm)	8.90 – 9.90	9.24
Osmolality (mOsmol/kg H ₂ O)	225 – 245	238
Background:		
WBC (K/μl)	<0.2	<0.2
RBC (M/μl)	<0.02	<0.02
HGB (g/dl)	<0.1	<0.1
PLT (K/μl)	<5	<5
Functional test	correct	correct

We hereby certify that this lot has been tested and was found to meet our specifications.
The production is in compliance to Quality Management System ISO 13485:2003.

Sign.: Anna Spryszak

Quality Assurance Department

This certificate is a computer print out and has therefore not been signed by hand.

PRINTED VERSION IS NOT SUPERVISED, CHECK TOPICALITY OF DOCUMENT BEFORE USE

Document number	Document version	Date of revision	Prepared by	Approved by
FOR048	2	10.01.2016	RW	JP