

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME: VARICELLA IgM CHORUS

Code 81079	
Lot No. 433	2019-08 2018-05

REAGENTS

Label	Description	Lot.	Exp	No
D2481	VARICELLA IgM	433	2019-08	6x6
B2481	CALIBRATORE VARICELLA IgM	433	2019-08	1
C2481	CONTROL SERUM VARICELLA IgM	433	2019-08	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	239	≥ 100
CONTROL SERUM (Index)	3,9	2,7-5,0
NEGATIVE CONTROL (Index)	0,0	< 0,7

CALIBRATOR

1	3	4	4	3	3	7	0	0	0	0	0	2	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

1	3	4	4	3	3	7	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

[Signature]

DATE 30/05/2018

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.