

Certificate of Analysis

Product Name: Coag Control Level 3 (Abnormal)-Orig:USA
Product Number: 100597
Lot Number: 503700

Expiration Date: 11/30/2016
Storage: 2 - 8°C

Specification	Specification Limit	Result
Appearance	Yellow to buff colored solid	Complies
Appearance Reconstituted	Yellow to straw colored, clear to hazy, free of particulates when viewed against dark background	Complies
Solubility	Less than 2 minutes	Complies
Vial-to-Vial Variability	<= 4 %	2
Moisture	<= 3 %	1
Thromboplastin D	26.0 - 34.0 S	30.9
Thromboplastin D	Confidence Interval*	26.3-35.5 s
Thromboplastin D	Value % PT	13 %
Thromboplastin D	Confidence Interval*	10-15 %
APTT-XL	65 - 79 S	73
APTT-XL	Confidence Interval*	62-84 s
Kontakt	Report	61 S
Kontakt	Confidence Interval*	52-70 s
100 NIH U/mL Throm.	150 - 450 mg/dL	292
100 NIH U/mL Throm.	Confidence Interval*	248-335 mg/dL

For in-vitro diagnostic use.

* Calculated from recovered values. Actual values recovered in the user laboratory will depend on reagent lot number, instrument and technique in use. Each laboratory should establish ranges using instrument, reagent and technique in place.

All individual donor units comprising this lot of Coagulation Control Plasma have been tested and found nonreactive for HBsAg and negative for antibodies to HIV and HCV. As with all blood products, this product should be handled as potentially infectious biological material.



Quality Representative

Date: 2014-06-20