

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: HERPES SIMPLEX 1+2 IgG -CHORUS

Code 81020

Lot No. 338



2018-08



2017-05

REAGENTS

Label	Description	Lot.	Exp	No
D2211	Herpes simplex 1+2 IgG	338	2018-08	6X6
B2212	CALIBRATORE Herpes simplex 1+2 IgG	338	2018-08	1
C2211	CONTROL SERUM Herpes simplex 1+2 IgG	338	2018-08	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	350	≥ 100
Control Serum (INDEX)	3,5	2,5-4,7
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

0	0	4	3	3	8	7	0	0	0	0	0	2	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

0	0	4	3	3	8	7	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschielli

APPROVED RESPONSIBLE HEAD, Quality Control Rea

[Signature]

DATE 08-08-2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.