



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-01
KIT DESCRIPTION:	LIQUID PROTEIN CALIBRATOR SERIES
CATALOGUE NO:	IT2691
BATCH NO:	440973
EXPIRY DATE:	2019-09

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
PROTEIN CONTROL II / 463LPC	FERRITIN	188	150-226	172	ng/ml
PROTEIN CONTROL III / 455LPC	FERRITIN	279	223-335	254	ng/ml

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE MRB R J Fitzgerald, W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0904 00.