

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| DATE OF MANUFACTURE: | 2018-01 |
|----------------------|------------|
| KIT DESCRIPTION: | GLUCOSE |
| CATALOGUE NO: | GL8318 |
| BATCH NO: | 434630 |
| EXPIRY DATE: | 2019-11-28 |

| SAMPLE | TARGET | RANGE | RESULT | UNITS |
|--------------------|--------|-------------|--------|--------|
| LEVEL II 892UN | 6.45 | 5.48 – 7.42 | 5.93 | MMOL/L |
| LEVEL III 766UE | 16.1 | 13.7 – 18.5 | 15.4 | MMOL/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

errer Danuelly

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | I www.randox.com