

RANDOX

CERTIFICATE OF ANALYSIS


This is to certify that the following kits were tested by
Randox Testing Services and approved by the QC Department.

DATE OF MANUFACTURE: February 2014
KIT DESCRIPTION: Immunoassay Speciality I – Level 1
CATALOGUE NO: IAS3113
LOT NUMBERS: 1322EC
EXPIRY DATE: 2015-10
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
C-Peptide (pmol/l)	1322EC	374 – 624	499	497	
				473	
PTH (pmol/l)	1322EC	2.39 – 3.99	3.19	3.6	
				3.8	

QC TESTED BY:



Janet Lightbody

DATE: 28-Feb-14

RESULT (PASS/FAIL): PASS

QC APPROVED BY:

