

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

DATE OF MANUFACTURE: 2015-11
KIT DESCRIPTION: ALT
CATALOGUE NO: AL1205
BATCH NO: 397664
EXPIRY DATE: 2017-08

SAMPLE	TARGET	RANGE	RESULT
LEVEL II / 1730CH	181	148 - 214	177mg/dl
LEVEL III / 1732CH	225	185 - 265	224mg/dl
LINEARITY:	241mg/dl		

APPROVED BY:



QUALITY CONTROL ANALYST