DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: WASHING BUFFER - CHORUS

Code 83606

Lot No. 403-C **Expiry** 2018-11

REAGENTS

Label	Description	Lot.	Exp	No
1035	WASHING BUFFER	403-3	2018-11	4

CONTROLS

	Found results	Expected results
na	na	na
na	na	na
na	na	na

STICK LABEL na

CALIBRATOR na

Label checked with barcode reader **YES** ☐ Operator Signature .....

CONTROL SERUM na

Label checked with barcode reader **YES** ☐ Operator Signature .....

STICK LABEL na

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD,Quality Control R.....

DATE 01/08/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL! na
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.