

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-12
KIT DESCRIPTION:	Glucose GOD-PAP
CATALOGUE NO:	GL3815
BATCH NO:	399900
EXPIRY DATE:	2018-11-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II / 892UN	6.45	5.48-7.42	6.18	mmol/L
Level III / 766UE	16.10	13.70-18.50	15.94	mmol/L

APPROVED BY:

**QUALITY CONTROL REPRESENTATIVE**