RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

\$ 2000

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-06
KIT DESCRIPTION:	HDL Cholesterol Reagent
CATALOGUE NO:	CH3811
BATCH NO:	454111
EXPIRY DATE:	2019-07-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II/ 2532CH	1.25	1.06 – 1.44	1.26	mmol/L
Level III/ 2534CH	1.53	1.30 – 1.76	1.53	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | I www.randox.com