

Certificate of Analysis

Product Name: APTT-XL - Orig:USA
Product Number: 100402
Lot Number: 506589

Expiration Date: 04/30/2016
Storage: 2 - 8°C

Specification	Specification Limit	Result
Appearance	Clear, yellow solution free of particulate matter.	Complies
Vial-to-vial Variability (CV)	$\leq 1.5 \%$	0.7
Level 1 Master Lot*	Lower Limit	26 S
Level 1 Master Lot*	Upper Limit	34 S
Level 1 Master Lot	Result	33 S
Level 1 Master Lot	Lot Number	147251
Level 2 Master Lot*	Lower Limit	42 S
Level 2 Master Lot*	Upper Limit	56 S
Level 2 Master Lot	Result	53 S
Level 2 Master Lot	Lot Number	251055
Level 3 Master Lot*	Lower Limit	62 S
Level 3 Master Lot*	Upper Limit	79 S
Level 3 Master Lot	Result	75 S
Level 3 Master Lot	Lot Number	210194
Heparin Level 2*	Lower Limit	76 S
Heparin Level 2*	Upper Limit	90 S
Heparin Level 2	Result	87 S
Heparin Level 2	Lot Number	406500
Normal Donors (Mean of 3)	Report	29 S
20% Factor VIII / Level 1 Control	≥ 1.10	1.31
20% Factor IX / Level 1 Control	≥ 1.07	1.18
20% Factor XI / Level 1 Control	≥ 1.21	1.41

For in-vitro diagnostic use.

All clot time specifications are for the Sysmex CA1500. They may be different for other photo-optical instruments.

* Control specifications are lot-specific. Before the stated control lots expire or are depleted, new ranges will be established for the replacement lots.



Quality Representative

Date: 2014-06-18