

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: a-TG - CHORUS

Code 86072	
Lot No. 280	Expiry 2017-10

REAGENTS

Label	Description	Lot.	Exp	No
D3571	a-TG	280	2017-10	6X6
B3571	CALIBRATORE a-TG	280	2017-10	1
C3571	CONTROL SERUM a-TG	280	2017-10	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	319	≥ 100
Control Serum (IU/mL)	433,2	222,6-676,9

CALIBRATOR

3	0	0	2	8	0	6	5	2	7	4	0	3	8	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *AS*



CONTROL SERUM

3	0	0	2	8	0	6	1	8	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *AS*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci
APPROVED RESPONSIBLE HEAD, Quality Control R *Benocci*
DATE 27/10/2016

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.