

Certificate of Analysis

Product Name: Coag Control Level 1 (Normal) - Orig:USA
Product Number: 100595
Lot Number: 151819

Expiration Date: 11/30/2019
Storage: 2 - 8°C

Specification	Specification Limit	Result
Appearance	Yellow to buff colored solid	Complies
Appearance Reconstituted	Yellow to straw colored, clear to hazy, free of particulates when viewed against dark background	Complies
Solubility	Less than 2 minutes	Complies
Vial-to-vial Variability	<= 1.8 %	0.9
Moisture	<= 3 %	1
Thromboplastin DS	Report	14.8 S
Thromboplastin DS	Confidence Interval*	12.6-17.0 S
Thromboplastin DS	Value % PT	76 %
Thromboplastin DS	Confidence Interval*	61-91 %
Thromboplastin D	Lower Limit	12.0 S
Thromboplastin D	Upper Limit	13.7 S
Thromboplastin D	Result	13.2 S
Thromboplastin D	Confidence Interval*	11.2-15.2 S
Thromboplastin D	Value % PT	72 %
Thromboplastin D	Confidence Interval*	58-86 %
APTT-XL	Lower Limit	28 S
APTT-XL	Upper Limit	35 S
APTT-XL	Result	32 S
APTT-XL	Confidence Interval*	27.4-37.0 S
Kontakt	Report	30 S
Kontakt	Confidence Interval*	25.8-35.0 S
Thrombin Time	Report	13.8 S
Thrombin Time	Confidence Interval*	11.0-16.6 S
Fibrinogen**	200 - 400 mg/dL	320
Fibrinogen**	Confidence Interval*	272-368 mg/dL

For in-vitro diagnostic use.

* Calculated from recovered values. Actual values recovered in the user laboratory will depend on reagent lot number, instrument and technique in use. Each laboratory should establish ranges using instrument, reagent and technique in place.

** Clauss fibrinogen

All individual donor units comprising this lot of Coagulation Control Plasma have been tested and found nonreactive for HBsAg and negative for antibodies to HIV and HCV. As with all blood products, this product should be handled as potentially infectious biological material.



Quality Representative

Date: 2017-06-12