DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: HERPES SIMPLEX 1+2 IgM -CHORUS

Code 81021

Lot No. 310 **Expiry** 2018-02

REAGENTS

Label	Description	Lot.	Exp	No
D2221	Herpes simplex 1+2 IgM	310	2018-02	6X6
B2221	CALIBRATORE Herpes simplex 1+2 IgM	310	2018-02	1
C2221	CONTROL SERUM Herpes simplex 1+2 IgM	310	2018-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	240	≥ 100
Control Serum (INDEX)	2,6	2,4-4,8
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

1	5	6	3	1	0	6	0	0	0	0	0	3	0	0	0	0	0

Label checked with barcode reader **YES** ☒ Operator Signature ....AB....



CONTROL SERUM

1	5	6	3	1	0	6	2	4	4	8	0	0	0	0	0	0	1

Label checked with barcode reader **YES** ☒ Operator Signature ....AB....

The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli

APPROVED RESPONSIBLE HEAD,Quality Control Reagents.....

DATE 06-03-2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.