

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	MICROALBUMIN
CATALOGUE NO:	MA2426
BATCH NO:	433092
EXPIRY DATE:	2018-07

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 922MA	32.3	25.8-38.7	30.1	mg/L
LEVEL II / 923MA	159.2	127.4-191.1	158.7	Mg/L

APPROVED BY:

4. Pandicani

QUALITY CONTROL REPRESENTATIVE