

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-07
KIT DESCRIPTION:	Cholesterol Reagent
CATALOGUE NO:	CH3810
BATCH NO:	452756
EXPIRY DATE:	2020-06-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II/ 1237UN	4.13	3.59 – 4.67	4.27	mmol/L
Level III/ 766UE	6.68	5.81 – 7.55	7.01	mmol/L

APPROVED BY:

**QUALITY CONTROL REPRESENTATIVE**