

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: April 2017
KIT DESCRIPTION: Immunoassay Premium Level 3
CATALOGUE NO: IA2640
LOT NUMBERS: 1668EC
EXPIRY DATE: 2019-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	1668EC	146 - 218	182	185	N/A
				182	
FT3 (pmol/l)	1668EC	12.6 - 21.0	16.8	16.4	N/A
				17.0	
CEA (ug/l)	1668EC	71.9 - 108	89.9	87.5	N/A
				90.0	

QC TESTED BY: Davina Graham
DATE: 13 Apr 2017
RESULT (PASS/FAIL): PASS
QC APPROVED BY: Adele Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15728. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016