



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |                     |
|----------------------|---------------------|
| DATE OF MANUFACTURE: | 2016-07             |
| KIT DESCRIPTION:     | CSF CONTROL LEVEL 3 |
| CATALOGUE NO:        | CF1501              |
| BATCH NO:            | 086CF               |
| EXPIRY DATE:         | 2018-10             |

| ANALYTE  | TARGET | RANGE     | RESULT | UNITS  |
|----------|--------|-----------|--------|--------|
| CHLORIDE | 112    | 103-121   | 105    | mmol/l |
| GLUCOSE  | 5.88   | 5.00-6.76 | 5.84   | mmol/l |

APPROVED BY:

Gráinne Toal

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5738, VAT Number GB 353 0304 00.

4706-QC Rev4  
27<sup>th</sup> September 2013