

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME: VARICELLA IgM CHORUS

Code 81079		
Lot No. 328	2018-07	2017-04

REAGENTS

Label	Description	Lot.	Exp	No
D2481	VARICELLA IgM	328	2018-07	6x6
B2481	CALIBRATORE VARICELLA IgM	328	2018-07	1
C2481	CONTROL SERUM VARICELLA IgM	328	2018-07	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	282	≥ 100
CONTROL SERUM (Index)	3,3	2,0-4,0
NEGATIVE CONTROL (Index)	0,0	< 0,7

CALIBRATOR

1	3	4	3	2	8	7	0	0	0	0	0	3	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature

13432870000360000
Varicella IgM
CALIBRATOR 0.175 mL IVD
LOT 328
2018-07 2°C-8°C
DIESE DIAGNOSTICA Monteriggioni - Italy CE

CONTROL SERUM

1	3	4	3	2	8	7	2	0	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature

1343287204000000001
Varicella IgM
CONTROL+ 0.425 mL IVD
LOT 328
2018-07 2°C-8°C
DIESE DIAGNOSTICA Monteriggioni - Italy CE

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo
APPROVED RESPONSIBLE HEAD, Quality Control Reagents
DATE 23/06/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.