

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-05
KIT DESCRIPTION:	Total Bilirubin
CATALOGUE NO:	BR3859
BATCH NO:	411746
EXPIRY DATE:	2020-03-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II / 892UN	34.4	27.2 – 42.6	35.0	μmol/L
Level III / 766UE	87.0	68.7 – 105.0	89.7	μmol/L

**APPROVED BY:** 

QUALITY CONTROL REPRESENTATIVE