

Certificate of AnalysisProduct Name: FDP Assay Kit - Orig:USA
Product Number: 100650
Lot Number: 603378Expiration Date: 11/30/2015
Storage: 2 - 8°C

Description	Lot Number	Expiration Date
Latex Reagent	561391	2016-11-30
Positive Control	516878	2016-07-31
Negative Control	603377	2015-11-30
Sample Collection Tubes	538065	2016-04-30
Glycine Buffered Saline	420881	2015-12-31

Specification	Specification Limit	Result
Positive Control **(using dropper)	3+ to 4+	4+
Negative Control**(using dropper)	Negative	Negative
1 µg/mL Fibrinogen	1+ to 3+	1+

For in-vitro diagnostic use.

** Effective concentration decreases from 40 µg/mL due to approximate dropper delivery volume.

Latex Reagent: 1 x 5 mL
Positive Control: 1 x 2 mL
Negative Control: 1 x 2 mL
Sample Collection Tubes: 30 Tubes
Glycine Buffered Saline: 1 x 100 mL
Slides: 10
Stir Sticks: 100
Plastic Pipettes: 30

This kit contains bovine-sourced material.

Each unit of source material used in the preparation of the Positive Control has been tested by an FDA approved method and found nonreactive for the presence of Hepatitis B surface antigen (HBsAg) and negative for antibodies to HIV and HCV. However, no known test can offer assurance that products derived from human blood will not transmit hepatitis or other infectious disease. This product, like all materials of human origin, should be handled as potentially infectious biological material.



Quality Representative

Date: 2015-01-30