

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: Epstein-Barr VCA IgG- CHORUS

Code 81055	
Lot No. 991	Expiry 2015-02

REAGENTS

Label	Description	Lot.	Exp	No
D2341	Epstein-Barr VCA IgG	991	2015-02	6X6
B2342	CALIBRATORE Epstein-Barr VCA IgG	991	2015-02	1
C2341	CONTROL SERUM Epstein-Barr VCA IgG	991	2015-02	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	430	≥ 100
Control Serum (INDEX)	3,3	2,4-4,4
Negative Control (INDEX)	0,0	$<0,7$

CALIBRATOR


0	0	5	9	9	1	7	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


0	0	5	9	9	1	7	2	4	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Laura capitani

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 16-12-2013

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.