



CERTIFICATE OF CONFORMITY

Certificate for Lot Number: 54635

Product Code: TS/5-17

Product: Probact Amies clear (M40-A2) medium in tube. Includes polystyrene shaft viscose tip swab in peel pouch (Pack Size 100)

Date Of Manufacture: 12/2017 Expiry: 2019-12 Shelf Life: 24 months

This is to certify that the product detailed on this certificate conforms to Technical Service Consultants Ltd procedures and specifications and has performed the tests detailed below within documented parameters.

Sterilisation Reference, Isotron PLC Number

S12029570-1-1 - 6197983 : Dose 25.9 - 47.4 kGy

For Irradiated products batch release is controlled by the verification of the radiation dose throughout the load.

Autoclave Reference, Time Manufactured

For autoclaved products, either a temperature datalogger or visual sterilisation indicator are used to confirm that sterility was achieved.

Physical and Microbiological Assurance

Tests Performed	Results Obtained
Batch Release (Sterility)	Pass
Microscope & Tube Drop Test as per M40-A2	Pass
Organism Validation of media according to M40-A2	Pass
-	
-	

Overall Result: Pass

Certificate Issued by: Towela Ngoma Certificate authorised on 22.12.2017

Electronically generated on 13th March 2018 - valid without signature

Sent to (compras@labcarecolombia.com)