



CERTIFICATE OF CONFORMANCE

We, **77 Elektronika Kft.** hereby certify, that

LabStrip U11 Plus

Catalog No.: ANA-9901-1

Lot No.:	9828-6318
Released quantity:	270 vials
Manufacturing date (dd.mm.yyyy):	05.2017
Expiration date (dd.mm.yyyy):	05.2019

Parameters:		Measured results: (ok/nok)	
Specific gravity:	1,000-1,030	ok ☒	nok ☐
Leukocytes:	normal-500 Leukos./ μ l	ok ☒	nok ☐
Nitrite:	negative-positive	ok ☒	nok ☐
pH:	5-9	ok ☒	nok ☐
Blood:	negative-300 Erys./ μ l	ok ☒	nok ☐
Protein:	negative-500 mg/dl	ok ☒	nok ☐
Glucose:	normal-1000 mg/dl	ok ☒	nok ☐
Ascorbic Acid:	negative-++	ok ☒	nok ☐
Ketones:	negative-+++	ok ☒	nok ☐
Urobilinogen:	normal-12 mg/dl	ok ☒	nok ☐
Bilirubin:	negative-+++	ok ☒	nok ☐

were produced and controlled in accordance with the valid Device Master Records
and Quality System Records.



Signature
Zsuzsanna Kocsis
Head of Reagent Quality Control

07.02.2018

Date (dd.mm.yyyy)

Note: