DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CARDIOLIPIN-G - CHORUS

Code 86046

Lot No. 304 **Expiry** 2018-01

REAGENTS

Label	Description	Lot.	Exp	No
D3311	CARDIOLIPIN-G	304	2018-01	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	304	2018-01	1
C3311	CONTROL SERUM CARDIOLIPIN-G	304	2018-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	260	≥ 100
Control Serum (GPLU/mL)	46,6	28,7-62,7

CALIBRATOR

3	3	5	3	0	4	6	3	3	3	5	5	2	3	8	5	0	0

Label checked with barcode reader **YES** ☒ Operator Signature .....



CONTROL SERUM

3	3	5	3	0	4	6	3	1	5	7	0	0	0	0	0	0	1

Label checked with barcode reader **YES** ☒ Operator Signature .....

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD,Quality Control Reagents.....

DATE 13/02/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.