

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-04
KIT DESCRIPTION:	UREA (LIQUID)
CATALOGUE NO:	UR3825
BATCH NO:	443000
EXPIRY DATE:	2019-06-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 892UN	7.04	5.98-8.10	7.07	mmol/L
LEVEL III / 766UE	19.20	16.30-22.10	19.11	mmol/L

**APPROVED BY:** 

QUALITY CONTROL REPRESENTATIVE