

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-03
KIT DESCRIPTION:	Bilirubin
CATALOGUE NO:	BR411
BATCH NO:	404564
EXPIRY DATE:	2020-01-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Total Level II / 892UN	30.3	24.0-36.6	31.5	umol/l
Total Level III / 766UE	82.9	65.5-100.0	82.0	umol/l
Direct Level II / 892UN	22.2	17.5-26.9	21.5	umol/l
Direct Level III / 766UE	29.4	23.2-35.6	28.3	umol/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Oavid Richmond

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com