



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                             |                                   |
|-----------------------------|-----------------------------------|
| <b>DATE OF MANUFACTURE:</b> | <b>2017-11</b>                    |
| <b>KIT DESCRIPTION:</b>     | <b>Alkaline Phosphatase (ALP)</b> |
| <b>CATALOGUE NO:</b>        | <b>AP3802</b>                     |
| <b>BATCH NO:</b>            | <b>434009</b>                     |
| <b>EXPIRY DATE:</b>         | <b>2019-02-28</b>                 |

| <b>SAMPLE</b>              | <b>TARGET</b> | <b>RANGE</b>     | <b>RESULT</b> | <b>UNITS</b> |
|----------------------------|---------------|------------------|---------------|--------------|
| <b>LEVEL II<br/>892UN</b>  | <b>174</b>    | <b>148 – 200</b> | <b>183</b>    | <b>U/L</b>   |
| <b>LEVEL III<br/>766UE</b> | <b>374</b>    | <b>318 - 430</b> | <b>178</b>    | <b>U/L</b>   |

**APPROVED BY:**

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
**T** +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** [marketing@randox.com](mailto:marketing@randox.com) **I** [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FAE Man. R J Fitzgerald, W Fitzgerald CBE, R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738, VAT Number GB 353 0304 00.

4706-QC Rev4  
27<sup>th</sup> September 2013