

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: TOXOPLASMA IgM - CHORUS

Code 81041			
Lot No.	388	2018-11	2017-11

REAGENTS

Label	Description	Lot.	Exp	No
D2281	TOXOPLASMA IgM	388	2018-11	6X6
B2281	CALIBRATORE TOXOPLASMA IgM	388	2018-11	1
C2281	CONTROL SERUM TOXOPLASMA IGM	388	2018-11	1

CONTROLS


	Found results	Expected result
--	---------------	-----------------

Calibrator (mOD)	364	≥ 100
Control Serum (INDEX)	3,0	2,3-4,2
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

1	3	1	3	8	8	6	0	0	0	0	0	3	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	3	1	3	8	8	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 15/12/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.