

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-07
KIT DESCRIPTION:	ISE URINE DILUENT
CATALOGUE NO:	RX3950
BATCH NO:	424533
EXPIRY DATE:	2019-03

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 773UC	CHLORIDE	82.9	70.5-95.6	82.3	mmol/L
LEVEL III/ 778UC	CHLORIDE	250.0	213.0-288.0	245.0	mmol/L
LEVEL II/ 773UC	POTASSIUM	30.1	25.8-34.6	28.0	mmol/L
LEVEL III/ 778UC	POTASSIUM	127.0	108.0-146.0	120.3	mmol/L
LEVEL II/ 773UC	SODIUM	63.5	55.9-71.1	63.4	mmol/L
LEVEL III/ 778UC	SODIUM	196.0	172.0-220.0	194.0	mmol/L

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S. P. Fitzgerald BSc PhD DSc (Hons) FIAC MAn. R. J. Fitzgerald, W. Fitzgerald CBE, R. P. Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI05738, VAT Number GB 353 0301 00.