

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-07
KIT DESCRIPTION:	DIRECT BILIRUBIN
CATALOGUE NO:	BR3807
BATCH NO:	419245
EXPIRY DATE:	2019-06

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 892UN	22.7	17.9-27.5	20.9	μmol/l
LEVEL III/ 766UE	30	23.7-36.3	31.5	μmol/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE