

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: ENA-6 S - CHORUS

Code 86012		
Lot No. 327	2018-04	2017-04

REAGENTS

Label	Description	Lot	Exp	No
D3131	ENA-6 S	327	2018-04	6X6
B3131	CALIBRATORE ENA-6 S	327	2018-04	1
C3131	CONTROL SERUM ENA-S 6	327	2018-04	1

CONTROLS

Found results	Expected results
Calibrator (mOD)	340 ≥ 100
Control Serum (ndx)	3,7 2,3-4,3

CALIBRATOR


1	0	2	3	2	7	6	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

with barcode reader

YES



Operator Signature



CONTROL SERUM

1	0	2	3	2	7	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 02/05/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.