

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: HERPES SIMPLEX 1 IgG RECOMBINANT - CHORUS

Code	81023		
Lot No.	434	2019-05	2018-05

REAGENTS

Label	Description	Lot.	Exp	No
D8680	Herpes simplex 1 IgG Recombinant	434	2019-05	6X6
B8680	CALIBRATORE Herpes simplex 1 IgG Recombinant	434	2019-05	1
C8680	CONTROL SERUM Herpes simplex 1 IgG Recombinant	434	2019-05	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	317	≥100
CONTROL SERUM (INDEX)	3,2	2,8-5,2
NEGATIVE CONTROL (INDEX)	0,0	<0,7

CALIBRATOR

2	5	8	4	3	4	6	0	0	0	0	2	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	5	8	4	3	4	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 29/05/2018

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.