

CERTIFICATE OF CONFORMITY

Certificate for Lot Number: 41822

Product Code: TS/72-MX Product: Mixed Protect, Bulk in Polythene Bag (Pack Size 1000)

Date Of Manufacture: 04/2013 Expiry: 2015-03 Shelf Life: 24 months

This is to certify that the product detailed on this certificate conforms to Technical Service Consultants Ltd procedures and specifications and has performed the tests detailed below within documented parameters.

Sterilisation Reference, Isotron PLC Number

S11009292-1-1 - 97716 : Dose 26.3 - 35.8 kGy, S11010051-1-1 - 106602 : Dose 26.1 - 35.6 kGy,
S11010839-2-1 - 114311 : Dose 25.5 - 34.8 kGy, S11011909-2-1 - 123192 : Dose 26.8 - 36.5 kGy
For Irradiated products batch release is controlled by the verification of the radiation dose throughout the load.

Autoclave Reference, Time Manufactured

For autoclaved products, either a temperature datalogger or visual sterilisation indicator are used to confirm that sterility was achieved.

Physical and Microbiological Assurance

Tests Performed	Results Obtained
Sterility	Pass
Product Specification	Pass
Recovery of <i>S. pyogenes</i> ATCC 19615 on Chocolate Agar/CO ₂ /36±	Pass
Recovery of <i>H. Influenzae</i> ATCC 49247 on Chocolate Agar/CO ₂ /3	Pass

Overall Result: Pass

Certificate Issued by: Joanne Walsh Certificate authorised on 05.04.2013

Electronically generated on 23rd May 2013 - valid without signature

Sent to Miguel Rojas (compras@labcarecolombia.com)