

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

Contraction

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-06
KIT DESCRIPTION:	CRP CONTROL LEVEL 3
CATALOGUE NO:	CP2481
BATCH NO:	2579CP
EXPIRY DATE:	2019-09-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL III 2579CP	CRP	145.9	116.7 - 175.1	154.5	mg/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com