

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SS-A - CHORUS

Code 86018	
Lot No. 966	Expiry 2014-08

REAGENTS

Label	Description	Lot.	Exp	No
D3181	SS-A	966	2014-08	6X6
B3181	CALIBRATORE SS-A	966	2014-08	1
C3181	CONTROL SERUM SS-A	966	2014-08	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	370	≥ 100
Control Serum (AU/mL)	43,2	21,7 - 64,6

CALIBRATOR


1	9	3	9	6	6	6	4	1	9	3	3	3	6	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


1	9	3	9	6	6	6	1	4	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Andrea Bellavista

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE: 10/09/2013

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.