

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-11
KIT DESCRIPTION:	CREATININE
CATALOGUE NO:	CR8316
BATCH NO:	430043
EXPIRY DATE:	2020-10

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 892UN	121.0	97.2-145.0	125.7	μmol/L
LEVEL III / 766UE	320.0	256.0-384.0	325.4	μmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

lillen Mujeg