

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: March 2017
KIT DESCRIPTION: Immunoassay Premium
CATALOGUE NO: IA2639
LOT NUMBERS: 1626EC
EXPIRY DATE: April 2017
SYSTEM OF ANALYSIS: Siemens Immulite 2000XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FT3 (pmol/l)	1626EC	8.10 - 13.5	10.8	11.8	N/A
				10.9	
SHBG (nmol/l)	1626EC	31.4 - 47.2	39.3	35.6	N/A
				39.7	
TSH (mIU/l)	1626EC	1.86 - 2.78	2.32	2.4	N/A
				2.45	

QC TESTED BY: Davina Graham
DATE: 21 Mar 2017
RESULT (PASS/FAIL): PASS
QC APPROVED BY: Adele Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S. P. Fitzgerald BSc PhD DSc (Hons) FIAE Man. R. J. Fitzgerald, W. Fitzgerald OBE, R. P. Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016