

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-03
KIT DESCRIPTION:	Microalbumin Cal set
CATALOGUE NO:	MA1567
BATCH NO:	448551
EXPIRY DATE:	2019-11-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level I/ 932MA	mALB	33.6	26.9 – 40.3	32.7	mg/L
Level II/ 933MA	mALB	165.7	132.6 – 198.8	172.2	mg/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com