

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-11
KIT DESCRIPTION:	URIC ACID
CATALOGUE NO:	UA3824
BATCH NO:	428569
EXPIRY DATE:	2019-04-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 892UN	340	290-380	341	μmol/L
LEVEL III / 766UE	560	490-640	565	μmol/L

APPROVED BY:

Yvette Millar

QUALITY CONTROL REPRESENTATIVE