

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	URINARY PROTEIN
CATALOGUE NO:	UP1570
BATCH NO:	427373
EXPIRY DATE:	2018-10

SAMPLE	TARGET	RANGE	RESULT	UNITS
URINE LEVEL II 833UC	0.150	0.105-0.195	0.150	g/L
URINE LEVEL III 866UC	0.256	0.205-0.307	0.290	g/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com