



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-11
KIT DESCRIPTION:	TRIGLYCERIDES
CATALOGUE NO:	TR8332
BATCH NO:	397999
EXPIRY DATE:	2017-11-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 892UN	1.04	0.87-1.21	1.01	mmol/l
LEVEL III/ 766UE	3.00	2.52-3.48	2.93	mmol/l

APPROVED BY:

Gráinne Toal

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R F Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI05738, VAT Number GB 353 0904 00.