

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-08-29
KIT DESCRIPTION:	Creatinine Reagent
CATALOGUE NO:	CR3814
BATCH NO:	444236
EXPIRY DATE:	2021-04-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II/ 1237UN	133	106 – 160	130	μmol/L
Level III/ 766UE	320	256 – 384	308	μmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE