

# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

|                                     |
|-------------------------------------|
| <b>DATE OF MANUFACTURE: 2015-09</b> |
| <b>KIT DESCRIPTION: AMYLASE</b>     |
| <b>CATALOGUE NO: AY3805</b>         |
| <b>BATCH NO: 400698</b>             |
| <b>EXPIRY DATE: 2016-11</b>         |

| SAMPLE             | TARGET | RANGE  | RESULT |
|--------------------|--------|--------|--------|
| LEVEL II<br>482UN  | 37     | 30-44  | 39     |
| LEVEL III<br>355UE | 84     | 67-101 | 92     |
| LINEARITY:         | 690U/L |        |        |

APPROVED BY:

*Louise Gilmore*

QUALITY CONTROL ANALYST