



Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

Specifications Microorganism Name: Enterococcus casseliflavus Catalog Number: 0761 Lot Number: 761-45 Reference Number: ATCC® 700327™* Purity: Pure Passage from Reference: 3	Expiration Date: 2018/11/30 Release Information: Quality Control Technologist: Kieshia L Negen Release Date: 2017/1/19
--	---

Performance	
Macroscopic Features: Medium to large, circular, convex, yellow/gray, glistening, translucent, slightly alpha hemolytic.	Medium: SBAP
Microscopic Features: Gram positive cocci.	Method: Gram Stain (1)
ID System: Vitek GP (1) See attached ID System results document.	Other Features/ Challenges: Results (1) Catalase (3% Hydrogen Peroxide): negative  Amanda Kuperus Quality Control Manager AUTHORIZED SIGNATURE

Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

Note for Vitek®: Although the Vitek® panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods.

⚠ Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.



(*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.



(1) These tests are accredited to ISO/IEC 17025:2005.

TESTING CERT #2655.01

bioMerieux Customer: 1012555
System #: 521711

Laboratory Report

Printed Jan 19, 2017 11:49 CST
Printed by: kln

Isolate Group: 761 45-1

Bench: KN

Card Type: GP Testing Instrument: 000017919CE3 (15220)

Bionumber: 534303265773771

Organism Quantity:

Comments:	

Identification Information	Card: GP	Lot Number: 242399410	Expires: Nov 20, 2017 12:00 CST
	Completed: Jan 18, 2017 16:56 CST	Status: Final	Analysis Time: 5.75 hours
Selected Organism	97% Probability Enterococcus casseliflavus		
	Bionumber: 534303265773771		Confidence: Excellent identification
SRF Organism			
Analysis Organisms and Tests to Separate:			
Analysis Messages:			
Contraindicating Typical Biopattern(s)			
Enterococcus casseliflavus AMAN(24),			

Biochemical Details

2	AMY	+	4	PIPLC	-	5	dXYL	+	8	ADH1	+	9	BGAL	+	11	AGLU	-
13	APPA	-	14	CDEX	-	15	AspA	+	16	BGAR	+	17	AMAN	+	19	PHOS	-
20	LeuA	-	23	ProA	-	24	BGURr	-	25	AGAL	+	26	PyrA	+	27	BGUR	-
28	AlaA	-	29	TyrA	+	30	dSOR	-	31	URE	-	32	POLYB	+	37	dGAL	+
38	dRIB	+	39	ILATk	-	42	LAC	+	44	NAG	+	45	dMAL	+	46	BACI	+
47	NOVO	+	50	NC6.5	+	52	dMAN	+	53	dMNE	+	54	MBdG	+	56	PUL	-
57	dRAF	+	58	O129R	+	59	SAL	+	60	SAC	+	62	dTRE	+	63	ADH2s	+
64	OPTO	+															

Installed VITEK 2 Systems Version: 07.01

MIC Interpretation Guideline:

AES Parameter Set Name:

Therapeutic Interpretation Guideline:

AES Parameter Last Modified: