



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-03
KIT DESCRIPTION:	GLUCOSE OXIDASE
CATALOGUE NO:	GL8318
BATCH NO:	442286
EXPIRY DATE:	2020-02-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 892UN	6.45	5.48-7.42	6.17	mmol/L
LEVEL III / 766UE	16.10	13.70-18.50	15.95	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald W Fitzgerald CBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.