



Trinity Biotech

Certificate of Analysis

Lot Number: 7447

Expiration Date: 2019-05-31

Production Date: 2017-05-09

Product Name: Premier Wash Reagent

REF: 01-03-0098

Shelf-Life: 2-Years (*Unopened*)

Manufactured By: Trinity Biotech, Kansas City, Missouri, USA

Storage Reqs: Keep tightly closed and store in a cool, dry location at 2-28°C, 36-82°F.

Use: For use with the Premier Hb9210 HbA1c Analyzer. No substitutions or other uses are permitted.

Method of Analysis:

Visual Examination

Standard: A clear, colorless solution free from observable particulate matter

Result: A clear, colorless solution is free from observable particulate matter.
Meets specification.

Specific Gravity

Standard: 0.985 – 0.995 Nominal

Result: Meets specification

Approval

Quality Manager:

Date:

5/15/2017