



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Radox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-06
KIT DESCRIPTION:	URINE ELEVATED ASSAYED CONTROL
CATALOGUE NO:	AU2353
BATCH NO:	748UC
EXPIRY DATE:	2018-04

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL III / 748UC	CALCIUM	4.79	4.31-5.27	4.52	mmol/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

Lisa Reer.

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Microsc, W Fitzgerald OBE R F Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5739, VAT Number GB 353 0304 00.