



Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

Specifications Microorganism Name: Acinetobacter baumannii Catalog Number: 0119 Lot Number: 119-31 Reference Number: ATCC® BAA-747™* Purity: Pure Passage from Reference: 2	Expiration Date: 2019/9/30 Release Information: Quality Control Technologist: Tracy A Blenker Release Date: 2017/11/16
--	---

Performance	
Macroscopic Features: Large, circular, convex, entire edge, cream, glistening Microscopic Features: Short, fat gram negative rods, singly and in pairs	Medium: SBAP Method: Gram Stain (1)
ID System: Vitek GN (1) See attached ID System results document.	Other Features/ Challenges: Results (1) Oxidase (Kovacs): negative Growth at 41 C: positive Growth at 44 C: positive  Amanda Kuperus Quality Control Manager AUTHORIZED SIGNATURE

Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

Note for Vitek®: Although the Vitek® panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods.

⚠ Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.



(*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.

(1) These tests are accredited to ISO/IEC 17025:2005.



bioMerieux Customer: 1012555
System #: 521711

Laboratory Report

Printed Nov 2, 2017 11:57 CDT
Printed by: tab

Isolate Group: 119 31-1

Bench: TB

Card Type: GN Testing Instrument: 000017919CE3 (15220)

Bionumber: 0241010103500352

Organism Quantity:

Comments:	

Identification Information	Card: GN	Lot Number: 241399320	Expires: Nov 19, 2017 12:00 CST
	Completed: Nov 1, 2017 18:37 CDT	Status: Final	Analysis Time: 6.00 hours
Selected Organism	99% Probability Acinetobacter baumannii Bionumber: 0241010103500352 Confidence: Excellent identification		
SRF Organism			
Analysis Organisms and Tests to Separate: Acinetobacter baumannii complex Acinetobacter nosocomialis Acinetobacter pittii 44C(1),41C(99), Acinetobacter baumannii 44C(99),Arg.hydr.(1),41C(99), Acinetobacter calcoaceticus Arg.hydr.(99),41C(1),			
Analysis Messages:			
Contraindicating Typical Biopattern(s)			

bioMerieux Customer: 1012555
System #: 521711

Laboratory Report

Printed Nov 2, 2017 11:57 CDT
Printed by: tab

Isolate Group: 119 31-1

Bench: TB

Card Type: GN Testing Instrument: 000017919CE3 (15220)

Bionumber: 0241010103500352

Organism Quantity:

Biochemical Details																	
2	APPA	-	3	ADO	-	4	PyrA	-	5	IARL	-	7	dCEL	+	9	BGAL	-
10	H2S	-	11	BNAG	-	12	AGLTp	+	13	dGLU	+	14	GGT	-	15	OFF	-
17	BGLU	-	18	dMAL	-	19	dMAN	-	20	dMNE	+	21	BXYL	-	22	BAlap	-
23	ProA	-	26	LIP	-	27	PLE	-	29	TyrA	+	31	URE	-	32	dSOR	-
33	SAC	-	34	dTAG	-	35	dTRE	-	36	CIT	+	37	MNT	+	39	5KG	-
40	ILATk	+	41	AGLU	-	42	SUCT	+	43	NAGA	-	44	AGAL	-	45	PHOS	-
46	GlyA	-	47	ODC	-	48	LDC	-	53	IHISa	+	56	CMT	+	57	BGUR	-
58	O129R	+	59	GGAA	-	61	IMLTa	+	62	ELLM	-	64	ILATa	+			

Installed VITEK 2 Systems Version: 07.01

MIC Interpretation Guideline:

AES Parameter Set Name:

Therapeutic Interpretation Guideline:

AES Parameter Last Modified: